



Excelsys Technologies Ltd.  
27 Eastgate Drive  
Eastgate Business Park  
Little Island  
Co. Cork, Ireland

Tel: + 353 21 4354716  
Fax: + 353 21 4354864

## DECLARATION OF CONFORMITY

**Product Name** : Xsolo Series  
**Product Type** : Component Power Supply  
**Model Number** : Xab-cdefgh

where:

**X** = all Xsolo part numbers start with 'X'

**a** = S or any other alpha character to denote market  
S = standard

**b** = 1000 or 500

1000 = 1008W output  
500 = 504W output

**c** = 24, 36 or 48

24 = 24V product  
36 = 36V product  
48 = 48V product

**d** = N; P or any alphanumeric character used to denote output voltage

N = Nominal output voltage  
P = Preset output voltage

**e** = '-'; C; R or S

'-' = standard model  
C = Conformally Coated  
R = Ruggedised  
S = C + R

**f** = Any alphanumeric character describing customer internal wiring lengths. Where no internal wiring exists and Screw Terminal Barrier Block only is used f = 0.

**g** = 00; 01; 02; 03; 04; 05; 06; 07; 08; 09; 10 or 11

00 = no options  
01 = I2C/PMBus  
02 = OR-Ing Function  
03 = 1 + 2  
04 = Low leakage  
05 = 1 + 4  
06 = 2 + 4  
07 = 1 + 2 + 4  
08 = Industrial Grade Y Capacitors  
09 = 1 + 8  
10 = 2 + 8  
11 = 1 + 2 + 8

**h** = Optional. Any alphanumeric character. Logistic use only.

See Page 2.

Excelsys Technologies Ltd. confirms that this product conforms to:

**IEC 60601-1-2: 2014 (4<sup>th</sup> Edition)**

Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests

Testing was conducted by Compliance Engineering Ireland Ltd.

Place of Issue: Cork, Ireland  
Date of Issue: 1<sup>st</sup> July 2016

For and on behalf of  
Excelsys Technologies Limited



Dermot Flynn